

**UE DECLARATION OF COMPLIANCE**

issued under the sole responsibility of the manufacturer

Producer: **APLISENS S.A.**,  
**ul. Morelowa 7, 03-192 Warszawa**

I declare with full responsibility that:

The Level Switches type:

**ERH-01÷04-16 (-K), ERH-07÷09-16 (-K)**

**ERH-01÷04-16.1 (-K), ERH-07÷09-16.1 (-K)**

in all the versions these are in compliance with the requirements of Directives:

- **LVD - 2014/35/UE** dated February 26, 2014


The standard: PN-EN 61010-1:2011+A1:2019-04 was applied for evaluation.

- **ROHS – 2011/65/UE** dated June 8, 2011

The standard: PN-EN IEC 63000:2019-01 was applied for evaluation.

- **ATEX – 2014/34/UE** dated March 23, 2016

Intrinsically safe Ex are made on the plate and certificate number:

 II 1/2G c Ex de IIB T4 Ga/Gb  
KDB 04ATEX352  
04ATEX352/1.../6

The certificate (Module B) was issued by the Notary Public No. 1453, the Central Mining Institute, 40-166 Katowice, Plac Gwarków 1.

The following standards were applied for evaluation: PN-EN 60079-0:2013-03+A11:2014-03, PN-EN 60079-1:2014, PN-EN 60079-7:2010, PN-EN 60079-26:2007, PN-EN ISO 80079-36: 2016, PN-EN ISO 80079-37: 2016 (*updated on 14/12/2017*).

Quality Assurance Notification (Module D) has been issued by Notary Public No. 1453, Central Mining Institute, 40-166 Katowice, Plac Gwarków 1.

- **Marine Certificates**

- **LR** - Lloyd's Register no. LR21290377TA
- **DNV** - Det Norske Veritas no. TAA000007V Revision No:2
- **BV** - Bureau Veritas no. 23604/C0 BV
- **PRS** - Polski Rejestr Statków (*Polish Shipping Register*) no. TE/1070/883400/23

Ostrów Wielkopolski, 18.02.2026

Sebastian Moch  
APLISENS S.A. DIRECTOR  
DIVISION IN OSTRÓW WIELKOPOLSKI

